

STATE OF ALASKA

Department of Public Safety
Division of Administrative Services



Pipette Calibration

ITB 2024-1200-0075

Amendment One

August 8, 2023

This amendment is being issued to answer vendor questions.

Kelly Pahlau
Procurement Specialist 3
Phone: (907) 269-8493
Email: kelly.pahlau@alaska.gov

Vendor Question:

Question 1: In reference to the ITB, it states multiple volumes and multiple readings. Do you have certain requirement? Would 2 volumes with 4 readings be sufficient? As for the multi-channel pipettes, would a visual check on all channels along with two channel verification be sufficient?

Answer: The state requires 3 volumes with 4 readings each, per channel. All channels on a multi-channel need to be verified, so visual checks will not be accepted.
