



THE STATE
of **ALASKA**
GOVERNOR MIKE DUNLEAVY

**Department of Environmental
Conservation**

Division of Administrative Services

555 Cordova Street
Anchorage, Alaska 99501

DECDAProcurement@alaska.gov

RFQ # 18-623-24 – Addendum One

To: RFQ # 18-623-24
Interested Parties

Date: Wednesday, March 13, 2024

From: Bryant Trujillo – Procurement Specialist 3
Department of Environmental Conservation
Division of Administrative Services

Subject: RFQ # 18-623-24
Addendum One

The solicitation package for **Request for Quote (RFQ) 18-623-24 – Sulfur Dioxide Trace Level Analyzer** for the State of Alaska, Department of Environmental Conservation (DEC), Division of Air Quality is hereby clarified or changed as follows:

1. Questions submitted by an interested party and answers provided by the Division of Air Quality representatives.

Addendum One is hereby made part of the RFQ and comprises two (2) pages.

All other terms and conditions for this RFQ remain unchanged.

Issued by: Bryant Trujillo
Title: Procurement Specialist 3
Email: decdasprocurement@alaska.gov

Questions & Answers

❖ Question 1 –

➤ Does the Division require 24-inch or 26-inch size rack mounts?

▪ Answer 1 –

- The Division requires 24-inch size rack mounts.

❖ Question 2 –

➤ Does the Division need the 50-amp option for zero and span valving?

▪ Answer 2 –

- Zero and span valving are unnecessary, as the Division has a separate gas dilution calibrator.

The quotation documents require individual acknowledgment of all addenda to the drawings and/or specifications. This is a **mandatory requirement**, and any quotation received without acknowledgment of receipt of addenda may be classified as being a non-responsive quotation.

End of Addendum